

# Protocol of RXR-BLA HEK 293T Cell-based Assay for High-throughput Screening

**DOCUMENT:** RXR-BLA\_TOX21\_SLP\_Version1.0  
**TITLE:** Protocol of RXR-BLA HEK 293T Cell-based Assay for High-throughput Screening

**ASSAY REFERENCES:**

| Assay Target                            | Cell Lines | Species | Tissue of Origin       | Assay Readout           | Assay Provider | Toxicity Pathway |
|---|------------|---------|------------------------|-------------------------|----------------|------------------|
| Retinoid X Receptor alpha (Recombinant) | HEK 293T   | Human   | Embryonic kidney cells | Beta-lactamase reporter | Invitrogen     | NR signaling     |

**QUALITY CONTROL PRECAUTIONS:**

1. -Cell culture is maintained by passaging twice a week and should not reach more than 90% confluence
2. -The assay should be performed in black-clear bottom 1536 well plates, so the bottom of the plates should not be touched

**MATERIALS and INSTRUMENTS:**

| Supplies/Medium/Reagent                | Manufacturer | Vender/Catalog Number |
|--|--------------|-----------------------|
| -Phenol red-free DMEM                  | -Invitrogen  | -Invitrogen/21063     |
| -DMEM                                  | -Invitrogen  | -Invitrogen/10569     |
| -Dialyzed FBS                          | -Invitrogen  | -Invitrogen/26400     |
| -Charcoal stripped FBS                 | -Invitrogen  | -Invitrogen/12676     |
| -NEAA                                  | -Invitrogen  | -Invitrogen/11140     |
| -Sodium pyruvate                       | -Invitrogen  | -Invitrogen/11360     |
| -HEPES                                 | -Invitrogen  | -Invitrogen/15630     |
| -Penn-strep                            | -Invitrogen  | -Invitrogen/15140     |
| -Hygromycin B                          | -Invitrogen  | -Invitrogen/10687     |
| -Zeocin                                | -Invitrogen  | -Invitrogen/R25001    |
| -Recovery Cell Culture Freezing Medium | -Invitrogen  | -Invitrogen/12648     |
| -0.05% Trypsin-EDTA                    | -Invitrogen  | -Invitrogen/25300     |

|   |                              |                                    |
|---|------------------------------|------------------------------------|
| -Black-clear bottom 1536 well plates            | -Greiner                     | -Greiner/789092F                   |
| -9-cis-Retinoic acid (Agonist control compound) | -Enzo Life Sciences          | -Enzo Life Sciences/BML-GR101-0005 |
| -Multidrop COMBI                                | -Thermo Electron Corporation | -Thermo Electron Corporation       |
| -BioRAPTR FRD dispenser                         | -Beckman Coulter             | -Beckman Coulter                   |
| -LiveBLAzer B/G FRET substrate                  | -Invitrogen                  | -Invitrogen/K1028                  |
| -CellTiter-Glo(R) One Solution Assay            | -Promega                     | -Promega/G8462                     |
| -Envision Plate Reader                          | -Perkin Elmer                | -Perkin Elmer                      |
| -ViewLux Plate Reader                           | -Perkin Elmer                | -Perkin Elmer                      |

## PROCEDURE:

### 1. Cell handling:

#### 1.1. Media Required:

| Component                              | Growth Medium     | Assay Medium      | Thaw Medium       | Freezing Medium |
|--|-------------------|-------------------|-------------------|-----------------|
| -Phenol red-free DMEM                  | -                 | -98%              | -                 | -               |
| -DMEM                                  | -90%              | -                 | -90%              | -               |
| -Dialyzed FBS                          | -10%              | -                 | -10%              | -               |
| -Charcoal stripped FBS                 | -                 | -2%               | -                 | -               |
| -NEAA                                  | -0.1mM            | -0.1mM            | -0.1mM            | -               |
| -Sodium pyruvate                       | -1mM              | -1mM              | -1mM              | -               |
| -HEPES                                 | -25mM             | -                 | -25mM             | -               |
| -Penn-strep                            | -100U/ml-100ug/ml | -100U/ml-100ug/ml | -100U/ml-100ug/ml | -               |
| -Hygromycin B                          | -100ug/ml         | -                 | -                 | -               |
| -Zeocin                                | -100ug/ml         | -                 | -                 | -               |
| -Recovery Cell Culture Freezing Medium | -                 | -                 | -                 | -100%           |

#### 1.2. Thawing method

1.2.1 -1ml frozen cells of RXRalpha-bla were taken in pre-warmed 10ml of thaw medium for centrifuging

1.2.2 -2-3ml of the thaw medium is taken to resuspend the pellet

1.2.3 -The cells were seeded in T-75 flask at 2 million cells.

#### 1.3. Propagation method

- 1.3.1 -The cells are detached using 0.05% Trypsin  
 1.3.2 -The cells are further passaged at a density of 4-5 million cells per T-225 flask

## 2. Assay Protocol

- 2.1 -Rinse the cells with DPBS and detach them by using 0.05% Trypsin and centrifuge  
 2.2 -Resuspend the pellet with assay medium  
 2.3 -Plate the cells in black-clear bottom 1536 well plate at 2000/well/6uL through 8 tip of a plate dispenser (Multi drop)  
 2.4 -Incubate at 37C for 5hrs  
 2.5 -Transfer 23nL of the compounds from the library collection and positive control through Pintool  
 2.6 -Incubate at 37C for 16hrs  
 2.7 -Add 1uL of CCF4 dye using a single tip of a plate dispenser (Bioraptr)  
 2.8 -Incubate at room temperature for 2hrs  
 2.9 -Read the fluorescence intensity through Envision plate reader  
 2.10 -Then add 4uL of CellTiter-Glo reagent using a single tip of a plate dispenser (Bioraptr)  
 2.11 - Incubate at room temperature for 30 min  
 2.12 -Read the luminescence intensity through ViewLux plate reader

## 3. Assay Performance

| <b>RXR<math>\alpha</math>-bla<br/>(9-cis Retinoic acid;<br/>Agonist control)</b> | <b>Online Validation<br/>Agonist<br/>(Mean <math>\pm</math> SD)</b> |
|--|---|
| EC50   | 8.40 $\pm$ 2.30 nM<br>(n = 27)                                      |
| S/B  | 2.19 $\pm$ 0.09   |
| CV (%)   | 5.03 $\pm$ 0.31 *<br>(n = 18)                                       |
| Z'   | 0.49 $\pm$ 0.07   |

\*CV values shown represent average of all plates excluding top 3 compound concentration plates.